

## EG-Konformitätserklärung

Die Lichtquelle:

## EC Declaration of Conformity

The light source:

### KL 2500 LCD

erfüllt die Bestimmungen folgender Richtlinien des Rates der Europäischen Gemeinschaft:

- 2004/108/EG (EMV-Richtlinie)
- 2006/95/EG (Niederspannungs-Richtlinie)
- 2002/95/EG (RoHS)
- 93/42/EWG (Medizingeräte-Richtlinie)

Die Lichtquellen sind für den Betrieb mit einer Nennspannung von 120V bzw. 230V ausgelegt.

Die Übereinstimmung der Lichtquellen mit den wesentlichen Schutzanforderungen obiger Richtlinien wird durch die vollständige Einhaltung folgender Normen nachgewiesen:

is in conformity with the following European Directives:

- 2004/108/EC (EMC Directive)
- 2006/95/EC (Low Voltage Directive)
- 2002/95/EC (RoHS)
- 93/42/ECC (Medical Devices Directive)

The light sources are designed for use with a nominal voltage of 120V or 230V.

Full compliance with the standards listed below proves the conformity of the light sources with the essential protection requirements of the above mentioned EC directives:

- EN 55011:2009
- EN 61000-3-2:2006
- EN 61000-3-3:1995+A1:2001+A2:2005
- EN 61326-1:2006

- EN 61010-1:2001
- DIN EN 60601-1 (VDE 0750 Teil 1):1996-03  
EN 60601-1:1990 + A1:1993 + A2:1995
- DIN EN 60601-1-2 (VDE 0750 Teil 1-2) :2002-10  
EN 60601-1-2:2001

Nach EN 62471:2008 ist die KL 2500 LCD ein Beleuchtungssystem der

#### Risiko-Klasse 1

Die Übereinstimmung mit den deutschen Ausführungen (DIN EN) aller oben angegebenen europäischen Normen (EN) ist ebenfalls nachgewiesen.

Die Lichtquelle trägt das CE-Konformitätskennzeichen

Acc. to EN 62471:2008 the KL 2500 LCD is an illumination system of

#### Risk Class 1

Full compliance with German versions (DIN EN) of all European standards (EN) listed above is also proven.

The light source bears the CE conformity mark



Mainz, November 2011

ppa.

Dr. Burkhard Danielzik

Business Unit Lighting and Imaging  
Vice President Europe

i.V.

Ralf Teichmann

Business Unit Lighting and Imaging  
Regional Sales Manager Europe